



Model: EDN.366
 Accuracy: ≤ 0.5%, ≤ 0.25%, ≤ 0.1% F.S
 Wetted parts: all stainless steel
 Application: water level measurement

Pressure Transmitter Model 기본판매가격(원)

Model:	EDN.366	A4 : accuracy ≤ 0.5%	510,210
		A7 : accuracy ≤ 0.25%	751,900
		A9 : accuracy ≤ 0.1%	966,800

Scope Code Specifications 선택사항 추가금액(원)

Analog output signals	O1	: 4...20mA, 2-wire system	-
Electrical connection	U	: underwater plug with IP68, 5m standard	-
Pressure connection	M14	: End-cap	-

Pressure Ranges / bar EDN.366 water level measurement 의 측정 압력범위는 사용자 요구에 의해 정하여 제작 공급된다.
 - 최소 제작 압력측정 범위: ≥ 0...20 mbar
 - 최대 제작 압력측정 범위: ≤ 0...20 bar

※ **Information**

- EDN.366 제품은 기본적으로 사용자의 요구 사양에 따라 제작 공급되어지는 제품이다. 따라서, 케이블 길이는 기본 5m를 공급하며 추가 길이와 측정 압력 범위는 사용자의 요구에 의해 제작 공급한다.
- 케이블 길이는 사용자 요구 사양에 따라 길이가 정해지며, 케이블은 대기압 보상도선이 있는 특수 케이블로 제작된다. 완제품 제작 후에, 이미 공급된 케이블에 추가적으로 연결이 불가능하므로 최초에 전체 길이를 정확하게 판단되어야 한다.
- 물이나 기름(비중)의 수심 깊이 또는 탱크 높이를 고려하여 측정 압력 범위를 정해야 하며, 이와 함께 케이블의 길이는 최종 연결지점까지의 전체 길이를 고려되어야 한다.

추가선택사항

Operating temperature	T4	: -40...+125 °C	63,200
4-wires system for current	C4	: 4...20mA, 4-wire system (+Vcc, +signal, -Vcc, -signal)	11,600
Underwater plug / IP68	ACMW	: additional extension underwater cable /m	6,500
Certificates	DMCC	: DAHO manufacturing calibration card	<i>price included</i>
	KC	: KOLAS calibration certificate	<i>하단의 KOLAS 교정비용 참조</i>

국제공인 인증성적서_KOLAS

분류 번호	측정기명	교정 주기	교정료				
			표준, 종합 교정절차서 실비	기본교정료	표준소급료	VAT별도	
20412	압력변환기 / 전송기	12					
	1	압력변환기-절대압용 압력계		171,500	17,150	188,650	
	2	압력변환기-게이지압용 압력계		114,700	11,470	126,170	
	3	압력변환기-차압용 압력계		118,200	11,820	130,020	
	4	압력변환기-기압계용 압력계		114,700	11,470	126,170	
	5	압력변환기-고압용 압력계 (100MPa 이상인 계기)		100 MPa 이상인 계기	143,660	14,366	158,026
	6	압력변환기-미압용 압력계 (0.01 MPa 이하인 계기)		0.01 MPa 이하인 계기	124,100	12,410	136,510
7	압력변환기-동압용 압력계		실비				

※ 다호제작품 구매하면서 교정할 경우에 위 "KOLAS 국제공인 교정료" 에서 50% 할인적용합니다.
다만, 다른 할인조건과 중복할인이 적용되지 않습니다.